

## Niedrigste Revisionsrate

### The Swedish Knee Arthroplasty Registry

OA / TKA	n	p-value	RR	95% CI
AGC	14,653		ref.	
PFC-Sigma	16,241	0.45	0.94	0.81-1.10
<b>NexGen</b>	<b>8,458</b>	<b>&lt;0.01</b>	<b>0.43</b>	<b>0.34-0.56</b>
Duracon	6,921	0.41	0.92	0.77-1.12
F/S MIII	6,618	0.02	0.79	0.65-0.96
Kinemax	1,975	<0.01	1.55	1.23-1.95
Scan	1,192	0.07	1.31	0.97-1.77
PFC	1,059	<0.01	1.68	1.27-2.22
Profix	623	0.68	0.89	0.51-1.55
AMK	469	<0.01	1.85	1.28-2.68
Natural II	396	0.68	1.17	0.55-2.49
LCS	346	0.12	0.53	0.23-1.18
MillerGalante II	298	0.40	1.27	0.73-2.21
PFC mobile bearing	221	0.27	0.46	0.11-1.84
Triathlon	202	0.84	0.82	0.12-5.87
Vanguard	161	0.98	1.03	0.14-7.36
Other	518	0.05	1.53	1.00-2.36
Gender (male is ref.)		0.93	1	0.89-1.11
Age (per year)		<0.01	0.96	0.95-0.97
Year of op. (per year)		0.34	1.01	0.99-1.0